PREMATURE BABY INCUBATOR

A. V. PENDSE

INDUSTRIAL DESIGN CENTRE
INDIAN INSTITUTE OF TECHNOLOGY
BOMBAY
1975

The-DP-50

Design of premature baby

Premature Baby Incubator
Diploma Project
A. V. Pendse
Industrial Design Centre

DEALS MAN THE THEORY IN A TRACE SELECTION OF STREET

Westmanning that the Land Committee of the L

· Propher Year truly of the state

Design of Premature Baby Incubator

Diploma Project

Submitted in partial fulfilment of the requirements of the postgraduate Diploma in Industrial Design

By

A. V. Pendse

I. D. C. Library
L. L. T. Bombay

DP/I - 50 / 1975

Industrial Design Centre
Indian Institute of Technology
Bombay
1975

Approval Sheet

Diploma Project entitled 'Premature Baby Incubator'

by A. V. Pendse is approved for the Postgraduate Diplome in Industrial Design.

Chairman:

Preminers: . Rolder. . Masaum

Industrial Design Centre
Indian Institute of Technology
Bombay.

Diploma Project Report

Project:

Design of Premature Baby Incubator

A. V. Pendse

Guide:

Shri U. A. Athavankar

Co-guide:

Shri M. Chattopadhyay

My acknowledgments to

Prof. S. Nadkarni
Shri U. A. Athvankar
Shri M. Chattopadhyay
Shri A. G. Rao
Dr. S. N. Deshpande
and other doctors
Technical staff
and my friends.

Contents:

Problem Statement	1
Introduction	2
Information	4
Analysia	18
Hypothesis	35
Synthesis	43
Photographs	48
Bibliography	50

Problem Statement:

To redesign an incubator with a particular emphasis on the following points:

- 1. Formal Appearance.
- 2. Hase of operation
- 3. Beenomy.

1. Introduction

In fact there are number of Indian and foreign incubators which are now available in India. Still, I feel that there is a need for redesign.

Initially, let us try to understand what is an incubator and why it is used in case of a premature baby. As the name suggests an incubator is essentially a space encloser with a facility of maintaining desired artificial environment inside that encloser.

A premature beby indubator artifically maintains womb temperature and provides a safty from infection to a prematurely born infent.

In the last few years manufacturers in the foreign countries are producing more and more sophesticated and advanced products, but in India first incubator was produced about five years before and all other Indian manufacturers copied the same design. There is no further development that has been done neither a study was done of an incubator to make it most suitable for Indian climate and situation. This incubator at present costs a lot and is beyond the reach of the majority of the hospitals.

Recently, Dr. S.N.Deshpande, a Civil Surgeon from Kolhapur studied this problem and developed an economical incubator

2. Information

The requisit information is collected by conducting a survey of hospitals having premature baby units, by visiting manufacturers, users, and through medical literature. This information is grouped under five sections such as:

- 1. Premature baby
- 2. Management and cure in the incubator
- 3. Present incubators in India
- 4. Socio-economic problems
- 5. Indian climate.

2.1.0 Premature baby.

2.1.1. Causes of premature birth (Low birth weight).

There are many factors which are considered to be associated with prematurity but the predisposing causes in India include -

- Socio-economic conditions such as lower class, poor housing mal-nutrition, negligence during pregnancy etc.
- Biological factors such as race small size of mother and birth order etc.



Comparison between Mature and Premature Infants The following table sets out the chief differences which can be observed on examination:

	Full Term	Premature		
Length Weight Proportion	20 in. (50 cm.) 7-8 lb. (3,180-3,640 g.) Circumference of head 14 in. (35°5 cm.) Circumference of chest 13 in. (33 cm.) Umbilicus midway between symphysis pubis and xiphi-	Less than 18½ in. (47 cm.). 5½ lb. (2,500 g.) or less. (Less than 13 in. (33 cm.) Less than 11½ in. (29 cm.) Umbilicus nearer symphysis		
Vitality	sternum. Strong and active Wakes for feeds Lusty cry Normal temperature Suction strong Pink and smooth	Weak and sluggish Drowsy Weak mewing cry Subnormal temperature Suction feeble or absent Red and wrinkled		
Nails	Subcutaneous fat present Nipples raised Hard Project beyond finger tips	Little fat present Nipples flat Soft Just to finger tips or not quite to finger tips.		
Genitals	Testicles in scrotum Labia minora covered by labia majora	Testicles in abdomen, inguinal canal or scrotum Labia minora not covered by labia majora		

I.D. C Library

As the majority of the Indian population comes under more or less all the factors mentioned above, rate of premature birth incidence is considerably high as compared to the developed countries.

Still there is a lack of a indigenous medical equipment which is within a reach of majority of hospitals.

2.1.2. Gauses of mortality

There are innumerable causes which leads to very high mortality rate, but majority of the incidents occure due to -

Infection
respiratory trouble
heat loss

which can be reduced by keeping an infent
in the incubator which provides

protection from infection
help in maintaining respiration
artificial desired temperature.

2.2.0 Management and care

Management and care in the incubator is based on the ordinary management of any new born baby but with certain modifications which are necessary to meet the special needs due to low birth weight.

The most important of these are :


- 1. Maintenance of body heat
- 2. Provision of humidity
- 3. Establishment of respiration and ventilation
- 4. Protection from infection
- 5. Careful observation and handling

consideration with all their access of

and the proper all the proper

- 6. Good management of feeding
- 7. Medical care.

2.2.1. Weintenance of body heat.



But it is more important to obtain stability of the infant temperature than to aim at particular temperature, and experience has shown that there is no need for anxiety unless the rectal temperature falls below 35°C or rises above 37°C. At 32°C the smallest and most weak infant tends to stabilize at 35 - 35°C.

Attempt to raise the body temperature to the supposedly normal leval of 37°C were unnecessary and might even provide detrimental to proper growth, as young premature babies have a thermo regulatory system adjusted to maintain a lower body temperature 35 - 36°C.



In the light of present knowledge incubator should be kept at a temperature of 32°C until the infants rectal temperature stabilizes about 37°C. Then the temperature of incubator can be reduced.

In few cases the temperature may have to be increased above 35°C but great care must be taken not to overheat the beby.

It is found out that infants loose 80% of body heat by radiation. Therefore heat supply by radiation is ideal.

Some investigators have maintained the skin temperature by infra red radiation but this method is too complicated, expensive and needs a very appheaticated controls.

2.2.2 Provision of humidity.

Doctors have found out that the rate of mortality is reduced by maintaining relatively high humidity inside the incubator. 65 - 70% humadity is considered optimum while higher humadity is maintained for smaller babies during early life as it maintains higher body demperature and have lower mortality rate at (80 - 90%) humadity.


2.2.3 Establishment of respiration and ventilation.

Respiration in smaller infants tends to be irregular in rythm and depth. Cough reflex is also absent making a danger of inhalation of regurgated fluids.

Such a danger can be reduced by keeping that infant in a tilted position.

Many times nasal passages are extremely narrow and the mucus membrance extremely weak. In such a case artificial respiration is maintained by an oxygen dome by closing an incubator lid. In such a case nasal route is never employed for administration of oxygen.


In case of a low weight infant without any complications, ventilation is maintained at the rate of 10 to 15 air changes per hour.



2.2.4 Protection from infection.

Premature infants are suceptible to infection. So it is isolated from others to all possible extent. Some incubators supply filtered air under slightly higher pressure that atmospheric pressure to prevent anything coming in from outside. For a precaution barrier nursing is strictly adeheard.

Additionally, premature baby unit is separated from other wards.




2.2.5. Careful observation and handling.

Incubator provides a clear view of an infant. As in premature infant any change in its body condition can be immediately recognised by observation. This can also helps in feeding, changing, examining etc.




Incubator should provide a facility
so that an infant can be nursed
without taking out of incubator.
These infants being extremely weak
should be handled as less as possible.



2.2.6. Management of feeding.

Incubator needs a claborate system
of feeding including a facility of
tilting. If sucking reflex of such
an infant is not developed, nasal
feeding is done which requires tilting.
Some times throbing is done.

2.2.7. Medical care.



As far as possible all the treatment of an infant should be done in the incubator. It should also provide head up or head down position for infant (± 15°) which is an essential nursing requirement.

I. D. C. Library

2.3.0. Present incubetors :

Present incubator is a development of open heated cot and is divided into two types on the basis of various factors such as type of nursery nursing procedure, use etc. Though from the structural and functional point of view the difference is very less.

Open Cot

Open Type

Closed Type

1. Type of nursary: Warm & humid

2. 21°0 60-65% H.

2. Less is the number of infants better it is.

3. 4.5 Sq.M/Cot 1.5 N c/e

4. OR Partition Sersen with .6M clear from all the sides.

5. Mainly for bigger infents if inc. is not available.

Hot & not humidified

Less is the number of infants better it is.

Meximum number 6 - larger infants

4.5 Sq.M/Inc.

10-12 A.C./hour Optimum humidity 65-70%

Most useful for larger infants. requiring ozygen but no other complications.

Warm & not humidified

Less is the number of infants better it is.

Meximum number 4 - Smeller infants.

2.67 Sq. M/ine.

1.5 M c/c

10-12 A.C./hour Optimum humidity 65-70%

Closed Type 6. Advantages : Cheap Various makes Various makes Can be heated Supply correct heat Supply correct heat easily. and humidity and humidity. Basy Og giving. Basy handling of Hasy observation Easy observation infant from distance from distance Freedom of access Isolates baby from to the infant air born infection. Ventilation with clear air (filtered) under pressure. Port holes & sleeves for care Facility of tilt-Facility of tilting Protection well worth ing. extra cost. 7. Disadvantages: Cannot give corre-Expensive Very expensive. heat & humidity Do not isolate baby Port holes and sleeves from air born infectend to obscure view. Og giving is diffition. oult. Infection.

Open Type

8. Westing system:

Open Cot

hot bottles or Electric blanket

9. No control

Heater for heating. Air.

Thermostat or other sophesticated controle with pilot lamp and Alarm.

Krater

Thermostat or other sophesticated controls with pilot lamp and Alarm.

2.4.0. Socio-Economic problems.

In India socio-economic problems dominate a product failure if not a design and therefore needs a careful attention for redesigning.

In India very few Government hospitals have premature baby unit. These units do not accept infants below 1500 gms. and such infants are left to God.

The information collected from doctors shows that more premature infants die in unit than outside.

Generally units are always over-loaded and under high pressure of work. So there is always a comparison in offering any facility to infant. Those who have batter chances of living are always preferred.

Negligence takes place due to high pressure of work and therefore barrier nursing is not followed very strictly.

Big city hospitals face peculier

problems of the staff. These hospitals do not get adequate number of nurses. Whichever nurse they train for premature unit, rashes up for foreigh country. While untrained nurse is incapable of handling a sophesticated incubator.

In the present Government hospital set up, doctors rarely have a chance for qualitative selection for equipment. So they tend to select lesser of the evil. Therefore, infants in the present incubator are likely to be burnt alive or cooled to death.

Present incubators are costly and functionally inadequate and being a copy of foreigh incubator, does not make use of favourable Indian climate. So nursing in this incubator also becomes costly.

Naturally within last five years of use of such an incubator. Indian doctors have developed an adverse openion about it.

2.5.0. Indian Climate

Climatic condition of various places in India throughout the year was studied from a critical angle and it is ascertained that Indian climate is favourable for nursing and this fact can be utilised in designing an incubator.

Statistical information gives the following result:


% Area under different zone at different period.

	Condition	Tempe .	. Humidity	JAN	MAY	AUG	MOA	AV A	TOT.
1.	Cold & dry	15°C	25-50%	2%	1%				
	Cold & humid		50-75%	12%	0.5%		5%		7%
	Cold & very hunid		75-100%	4%	0.5%	1%	4%		
2.	Cobl & dry	15-2000	25-50%	13%	2%		5%		
	Cool & humid		50-75%	23%	2%	2%	8%		16%
	Cool & very humid		75-100%	2%	1%	4%	3%		
3.	Temperate & dry	20-30°C	25-50%	17%	4%		19%		
	Temperate & humid	141	50-75%	23%	15%	28%	45%	weight.	60%
	Temperate & very humid		75-100%	4%	10%	60%	12%		
4.	Hot & very dry	30-35°C	25-50%		8%				
	Hot & dry		50-75%		35%		· ·	O	17%
	Hot & humic		75-100%		19%	5%		٠	
	Hot & very humid				1%				
5.	Very hot & dry				1%	Ъ			

This shows that 76%n places in India have temperature between 15-30°C while more than 80% places have humidity in between 50-75%.

3 Analysis.

3.1.0. Structural analysis.



An incubator is made up of two principal units, the infant chamber and heating chamber which are created by putting a mesh partition in between.

The total assembly comprises of M.S. angles fitted with glass panels with a hinge operated lid at the top. It is finished with M.S. sheet.

'L' shape M.S. angles which are real structural members seems to be bigger and stronger than the necessity. 6 mm thick plain or toughened glass satisfy the structural requirement.

M.S. angles and sheet (18 or 20 gauge)
is welded together which makes the total



assembly stronge. In the open type, top lid serves as an access, while in the closed type, port holes and sleeves is an additional access.


The incubator is mounted on a steel cupboard with a stacking facility which also offers tilting facility. The incubator takes support of cupboard for tilting. The cupboard is fitted with rubber castors for mobility additionally. Incubator also has handles for carrying it in hands which looks delicate as compared to the size and weight of the incubator.


Systems.

Incubator has air inlet at the bottom and air outlet at the top both operatable from the front. Both are structurally very simple to manufacture and to operate.

Heating mechanism comprises of a heater along with a radiating sins. This system demands a substantial space.

A stainless steel water trough is mounted on the heater along with a water





filling funnel and draing out cock fitted from outside. Though structurally sound, the system is too complicated to assemble as it is liable to leak.

Control Panel :

Control panel is made out of M.S. sheet also acts as a support and access to heating element. It is well marked to indicate knob position.

Additionally a hygrometer is maintained inside the incubator which gives temperature and humidity inside.

Safety device :

Heater is associated with thermostats or other control devices, which are mostly imported, along with Alarm and pilot lamp system which is situated in a heating chamber or can be connected to the incubator.

It seems that the present indubator
design is dominated by the production
method. There are number of manufacturers but nome of them produces more than

100 pieces per year. So all the details are suitable for fabrication and assembly rather than a mass production.

3.2.0 Functional analysis

The basic function of an incubator is to maintain a prematurely born infant in it at a temperature within a range of 30° to 37° C, relative humadity of 60 to 90%, ventilation at the rate of 10 to 15 Airchanges per hour and give aprotection to the infant from infection. Present incubators have various systems to serve these needs.

3.2.1. Heating - Old system :

- a) Use of hot water bottles along the infant Though this system is the cheapest, this does not maintain a constant temperature. A constant supervision and frequent hot water replacement is needed. Leakage is an additional nuisance.
- b) In this system, air is passed around hot water drum, but this makes a system structurally complecated and the total incubator large in size.

Present system :

a) Use of heater - Use of heater made incubator extremely suitable for maintenance. It also reduced its unnecessary bulk.

Now a days heater is used with the following accessories for control.

Heater + thermostat - made erude and sturdy but no precision.

Heater + thermocouple - precise but very delicate.

Heater + timer - precise but costly

Heater * thermister - precise but very costly.

but such costly devices made incubator beyond the reach of majority of the hospitals. Where after-sale care of equipment is not envisaged, this proved to be a failure.

Heater of such a high capacity is used to reduce time of heating but in the Indian climate, where ambient temperature and required temperature is not much, it is unnecessary.

Incubators do heating in two ways.

- a) In some incubators heating chamber is located by the side of infart chamber. In this case, body portion nearer the heater gets more heat and heating becomes uneven.
- b) In this system, heater is located below the bed and heating is also even but there is an increased danger of an accident as heater is just below the infant.

Some manufacturers go for infra red light heating by radiation but this system is extremely complecated and therefore not welcomed by the hospitals.

All the present incubators have oblong shape. Therefore surface area of the incubator is more as compared to the volume of air in it. This increased heat loss by the air additionally air remains stagnant in the corner and cool down which results in uneven heating.


This proves that the present incubators shape is not functional in retaining heat.

Market Militaria da mare Moras

3.2.2. Humidity :

High humidity is maintained in a simple way by keeping a water trough inside the infant chamber but this method can not give required (between 60 to 90%) percentage.

In some incubators, water trough is kept above heater. Many times this creates excess vapour pressure. The required precise percentage is achieved by increasing or reducing effective contact of water surface with air but this system is too complecated and is unnecessary in Indian &p conditions. Water pouring and draining system is not very hygienic but reduces the work of nurse. Water trough is not seen from outside.





3.2.3 Ventilation :

Depending on the type of incubator, mainly two systems are followed:

Natural Flow System -

In this system, inlet is at the lowest level end outlet at the top. Air is passed on the heating element first and then on the infant. This system does not isolate infant in a strict sense and therefore degree of protection from infection is less.

Forced ventilation -

In this system, filtered air is passed on a water surface and heating element under pressure which is slightly more than atmospheric pressure. This reduces chances of infection from the inlet as well as out let as inside air prevents outside air entering from the out let.

Whenever oxygen is given, inlet is closed and oxygen is given through oxygen inlet (Oxygen 1: 3 Air)





3.2.4. Infant care.

Incubator provides the facility of observing infant from out side through the glass without opening it, but this do not offer a complete view. It also provides protection to the infant from any object falling on it. For this reason incubator should not be opened from the top.

Incubator bed provides a comfortable soft surface which can be tilted to ± 15° to maintain easy respiration or circulation.

In the open type incubator, infant is taken out of the incubator for nursing. Though, this saves lot of nursing time, is bad for infant health. Incubator design does not facilitate nursing inside incubator.

Closed type incubators have facility of nursing within incubator. Such incubators require more space for infant chamber.

Time required for nursing is also considerably more but it provides the best possible treatment to the infant.



3.2.5. Storage.

Though an essentiality as far as possible this facility is rejected. As present incubator cupboard is heavy and very big and hassubstantial height. This is under utilised and reduces mobility of the incubator cup-board, also facilitates tilting.

3.2.6. Control.

Present incubators have on-off switch
and two control knobs with a pilot
lemp but to explain a selection and
knob position, there is a completed
explanation sheet. This makes a simple
control difficult to operate.

3.2.7. Maintenance.

Present incubator design is most utilitarian and has beer minimum perts. This is easy to clean and maintain, but if the heating system fails, this incubator can not be repaired without the help of a skilled agency. 3.3.0. Formal analysis is restricted to only Indian incubator.

All the present incubators have the same form which differ only in detailing of transition from one plain to another.

The visual proportions a total unit is odd. It is very high and has less width.

All incubators avoid sharp corner but they can not hide deformation at the joint of threeplains.

Angles supporting glass panel gives a very rigid and harsh shape to glass panel.

Opaque frame work stands in constrast to transperent glass panel and is responsible for giving it a sturdy appearance.

Four verticle angles acting as legs do not have good proportions. They are more wide and short.

Bottom and rear side is true to its

structure without any surplus finishing.

Sides of the incubator are plain with number of parts stuck on it.

Oxygen inlet on one side is hardly recognisable.

Funnel for pouring water, and water draining cock have totally different character. The point is strictly ulitilarian sides are enamel finished while tap and funnel are out of stainless steel.

Stainless steel strips for tilting adjustment look as if unnecessary extension stuck on the side. It has a functional form.

Standard handles on both sides and on the top lid do not go with the total assembly. They look extremely delicate.

Control panel on the front is not well differentiated. It is in the same plain and of the same colour.

Knobs and pilot lamps are so much away from each other that they loose their identity as one unit of controls.

Knobs are small as compared to the front panel and black colour on grey or white back ground gives a still smaller impression.

Graphics is well marked.

Incubator is finished with grey or while enamel and has a smooth finish, which gives a total unit polished and sophesticated appearance.

3.4.0. Ergonomical analysis.

3.4.1. Infant space.

shape and infant chamber is around

24" x 18" x 18" (height). Considering

infant's body measurements at birth and

at the time of discharge, this chamber

does not allow rolling or fully streched


body positions.


Cotton mattress used in the Indian incubator is warm and confortable. In some foreign hospitals water beds are used but they are extremely cool. Thermocole ball mattress is soft, comfortable, quick drying, easy to clean and it allows air circulation within the mattress.


Present incubator beds can be tilled through ± 15° which fully satisfy all needs for nursing.

3.4.2. Nursing problems.

Present incubator being a copy of foreign design, does not take care of physical limitations due to body measurement of Indian nurses.







This incubator takes support of the trolley for tilting and the whole unit is tilted. This is tiresome and nurse tends to add support under the bed rather than tilting it.

In case of a present incubator, for any operation bending is essential.

The length of an incubator is such that one man can not lift it easily. It has only one handle on each side so in transport it may swing towards right or left.



- 4.0.0. Hypothesis.
- 4.1.0. Structural hypothesis.
- 4.1.1. Incubator should be developed with an idea of fabrication and assembly by manual work. Production process should avoid methods of mass production.
- 4.1.2. Incubator should be light in weight, strong against any shock. It should have simple joining details which makes the unit air tight.
- 4.1.3. It should protect that infant from any object falling on it.
- 4.1.4. It should provide a comfortable bed and an easy tilting arrangement (± 15°).
- 4.1.5. Incubator should be structurally simple to assemble and dismentle. It should be easy to clean. In case of a system failure, its parts must be easily replaceable without the help of any skilled agency.
- 4.1.6. Incubator should have water pouring facility, very simple and water level

should be seen from outside.

- 4.1.7. Structurally incubator should provide facilities such that the incubator can be utilised both ways, either as open type or as a closed type.
- 4.1.8. Incubator should be made up of two
 principal units. The infant chamber
 and heating chamber. Heating chamber
 should be located below the infant
 chamber.
- 6.1.9. Assembly should avoid 'skin and bone' construction. It should reduce number of parts to the minimum possible extent.
- 4.1.10. Heating element should avoid over heating and should be economical and easily replaceable.
- 4.1.11 Electrical risk should be avoided.

 Infant chamber should be well guarded against any power leakage and fire hazard.

- 4.2.0. Functional hypothesis.
- 4.2.1. The shape of the incubator should be such that it should hold maximum air volume in the minimum surface area.
- 4.2.2. Incubator should provide complete and clear view of the infant.
- 4.2.3. Incubator should avoid sharp comers.

 Infant chamber should provide heat by rediction to the maximum possible extent.

 It should also reduce heat loss by convection or conduction.
- 4.2.4. Heating element should be capable of providing any temperature within a range of 30 to 37°C with an accuracy of ± .5°C. Heating element should reach the required temperature within 15 minutes.
- 4.2.5. Heating must be even throughout the chember.
- 4.2.6. Incubator should maintain a water trough in such a way that it maintains a relative humidity of 60 to 90 %.


- 4.2.7. It should maintain a ventilation of
 10 to 15 air changes per hour, but it
 should not allow any air leakage i.e.
 air entering inside from out excepting
 the inlet for the air.
- 4.2.8. Incubator should isolate infant from other environment to all possible extent.

 It should offer an infant protection from the infection.
 - 4.2.0. Incubator should provide a facility of nursing within the incubator.
 - 4.2.10. Incubator bed should be soft and worm and of quick drying material. It should not offer strain to body muscles. The bed should have a facility of tilting it through ± 15°.
 - 4.2.11. Incubator control should be easy to operate and free from any reference or calculations.
 - 4.2.12. Incubator should be easy to clean, maintain and to repair.

- 4.5.0. Formal hypothesis.
- 4.3.1. Incubator should have a very simple and attractive form. The visual proportions of the total unit should be pleasant.

 The form should be truly functional.
- 4.3.2. It should be free from any deformation at the joint. It should avoid sharp corners or edges.
- 4.3.3. Transperent panel should have an attraetive shape which is functional and goes with the total unit.
- 4.3.4. The unit should have a smooth surface finished with opaque white or ivory colour from inside and white or grey or any colour from outside which gives it a dignity.
- 4.3.5. Any element or fixture on the outer skin of the incubator should have a convincing form with respect to its function and should be well merged with the total assembly.

- 4.3.6. If necessary, colour coding or other symbols should be used to indicate a certain function, priority or emergency.
- 4.3.7. Control panel should be in proportion to the incubator. It should be well differentiated and well marked on the front face. It should have very simple form finished with an attractive and inviting colour. It should be well marked or embossed with a graphics etc. It should be well organised.



4.4.0. Ergonomical hypothesis.

4.4.1. Infant space in the incubator should be adequate. It should allow rolling or free body movements within the chamber.



4.4.2. Infant chamber should also facilitate
free movement of the hand (and fore arm)
and easy reach for all the nursing
requirements.

4.4.3. It should use a mattress which is comfortable, easy to maintain, easy to clean and quick drying. It should remain warm and should allow air circulation within the mattress.

4.4.4. Incubator bed should have a facility of tilting it through ± 15°.

4.4.5. Incubator design should take care of physical limitations due to body measurements of Indian nurses being an operator of the equipment.

4.4.6. Incubator should be supported on a stand in such a way that incubator top is easily operatable.

- 4.4.7. Incubator bed level and opening should be such that the infant can be easily 'kept in' and 'taken out' of it without lifting it.
- 4.4.8. Incubator should be such that nursing can be done without leaning forward unnecessarily.
- 4.4.9. Incubator should be easy to lift or to transport by one person.
- 4.4.10. Control should be designed to give

 accuracy in manual adjustment. Knobs

 should be designed to have a firm grip

 and easy adjustment.
- 4.4.11 Graphics, colour code, letters (type)
 etc. should be carefully selected
 taking into consideration time required
 for setting and limitations of human
 eye and human action.

A survey of the existing incubators and the needs of the hospitals who use them showed that there is a need of an economical incubator to serve, mainly hospitals in the small town where skilled but untrained (for premature unit) nurses are available.

As incubators are manufactured on ordered by the process of fabrication and assembly, this leads to a design of details suitable for such industry only.

The main design constrains were complete view of the infant, maintenance of heat, ventilation, humidity and protection from infection which leads to an air tight and rigid construction which is easy to manufacture, maintain and to repair.

A number of shapes were sketched and studied in terms of their suitability. They were also visualized with different kinds of systems (alternatives).

Initially a rectangular shape was

L. D. C. Library

selected, but it was found out that this shape does not solve any functional problems, especially that of heating. This is uneconomical from the structural point of view though it is economical and easier to manufacture. It is not suitable from the ergonomical point of view.

So an oval shape was selected which solved many problems but raised constructional and manufacturing difficulties.

So a golden mean was reached by the selection of circular shape.

The final design has the following feature

Hood - It is made up of transperent polystyrene and has a domical shape which is formed by forming process. It is made into two halves which slide one above other. The outer portion serves as opening while inner portion has two holes which act as a restricted access to the infant for nursing. It also helps in retaining heat and radiating it on the infant by the virtue of its shape. It holds more air volume as compared to its





surface area. It helps in maintaining oxygen dome.

Ded - Incubator bed is made out of thermocole balls which is easy to clean and maintain. It is quick drying and allows air circulation within. It is warm and comfortable. It is circular in shape.

Bed-dish - is made out of aluminium sheet by forming method. It also facilitates tilting through ± 15 ° and takes its support on the base.

Base - Incubator base is made up of aluminium sheet by fabrication process. The joining is done by synthetic adhesive to make it air tight. It has a shape of a frustrum which has smaller diameter at the bottom. It supports the whole unit.



Heating system - heating system is situated exactly below the infant bed. It comprises of 4 bulbs as a heating element, the heating of which is controlled by a Rehostet by increasing or reducing the voltage. The whole system is situated in a aluminium box with an asbestos lining from inside.



Ventilation system - Incubator has an air inlet at the rear side of the base. The air is passed on the bulbs and enters in infant chamber along the periphery of the bad. It has an outlet in the incubator hood which is adjustable to give 10 to 15 air changes per hour.

portion of the base. It is sunked to give protection to the thermameter. The sliding indicator when matched with the mercury colimn gives the required temperature. Thermameter has seven positions to adjust. The thermameter when adjusted to give 30°C temperature any position of indicator which metches with the mercury column ensures inside temperature of 30°C. Same is true of any selected position of

a thermaneter for 32. 35 or 370 c etc.

Control panel is situated on the front



- 5.1.0. Salient features of the design.
 - Use of inexpensive and easily available materials.
 - 2. Easily replacable heating element.
 - 3. Construction permits easy changing of elements and also facilitates dismantling and cleaning it easily.
 - Provision of complete and clear infant
 view. Observation from a longer distance
 - 5. It is easy for nursing infant.
 - 6. Reduces the risk of over heating.
 - 7. Control is safe and simple.
 - 8. Filting is very easy.
 - 9. It avoids a sharp corners and edges.
 - 10. Use either as open type or closed type.
 - 11. Simple form which reduces production cost and make it better looking.
 - 12. It can adjust with other hospital equip-
 - 13. The visual bulk is reduced by the use of colour scheme and form.

· Bibliography

- 1. The Premature Baby
- 2. Prematurity
- 3. Incubator Design

V. M. Crosse.

Beryl Corner. -

K. K. Trivedi.

